



[1] **EC-TYPE EXAMINATION CERTIFICATE**  
according to Directive 94/9/EC, Annex III  
(Translation)

[2] Equipment and Protective Systems intended for use in Potentially Explosive Atmospheres, Directive 94/9/EC

[3] EC-Type Examination Certificate Number: **IBExU05ATEX1106 X**

[4] Equipment: Pressure measuring device type AX12 x|act nn...  
AX12 XMP nn...

[5] Manufacturer: BD SENSORS GmbH

[6] Address: BD-Sensors-Str. 1, D-95199 Thierstein

[7] The equipment mentioned in [4] and any acceptable variation thereto are specified in the schedule to this EC-Type Examination Certificate.

[8] IBExU Institut für Sicherheitstechnik GmbH, NOTIFIED BODY number 0637 in accordance with article 9 of the Council Directive 94/9/EC of 23<sup>rd</sup> March 1994, certifies that this equipment has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of equipment intended for use in potentially explosive atmospheres given in Annex II to the Directive.

The examination and test results are recorded in test report IB-05-3-190 of 3<sup>rd</sup> August 2005.

[9] Compliance with the Essential Health and Safety Requirements has been assured by compliance with EN 50014:1997+A1+A2, EN 50020:2002 and EN 50284:1999.

[10] If the sign "X" is placed after the certificate number, it indicates that the equipment is subject to special conditions for safe use specified under [17] in the schedule to this EC-Type Examination Certificate.

[11] This EC-Type Examination Certificate relates only to the design and construction of the specified equipment. If applicable, further requirements of this directive apply to the manufacture and supply of this equipment.

[12] The marking of the equipment mentioned in [4] shall include the following:

 **II 1G, II 1/2G or II 2G EEx ia IIC/IIB T4**  
-25 °C ≤ T<sub>a</sub> ≤ +70 °C

The validity of the marking for the type of protection goes by the marking table in test report.

IBExU Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - D-09599 Freiberg  
Tel.: +49 (0)3731 3805-0 - Fax: +49 (0)3731 23650

Authorised for certifications Explosion protection  
By order

(Dr. Lösch)



Freiberg, 3<sup>rd</sup> August 2005

Certificates without signature and seal are not valid.  
Certificates may only be duplicated completely and unchanged. In case of dispute, the German text shall prevail.

**Schedule**



Summary of the Test Result:

The Pressure measuring device type AX12 x|act nn... and AX12 XMP nn... fulfils the requirements of the type of protection Intrinsic Safety for an electrical apparatus for the Group II, depending on execution Explosion Group IIC or IIB and the Categories 1G, 1/2G or 2G.


[17] **Special conditions**

- The equipment designed with connector have to be installed in such a way, that the Degree of protection IP20, always will be kept.
- The safety and assembly notes contained in the operating instructions and the Ambient temperature range from -25 °C to +70 °C have to be observed.
- At pressure transmitter with the marking category 1/2 equipment, the sensor diaphragm serves as partition wall and has to be protected against mechanical damages.
- The isolation of the intrinsically circuit opposite the case is because of leakage flows in the blocking capacitors from the EMV-boards limited.

[18] **Essential Health and Safety Requirements**

Confirmed by norms (see [9]).

By order



(Dr. Lösch)

Freiberg, 3<sup>rd</sup> August 2005





[1] **1<sup>st</sup> Addition to  
EC-TYPE EXAMINATION CERTIFICATE IBExU05ATEX1106 X  
- Translation -**

[2] Equipment:  
a) Title corresponding to IBExU05ATEX1106 X:

Pressure measuring device type      AX12 x|act nn ...  
AX12 XMP nn ...

b) Supplementary titles: Pressure measuring device type      AX12 HMP331...

[3] Manufacturer:      BD SENSORS GmbH

[4] Address:      BD-Sensors-Str. 1  
95199 Thierstein  
GERMANY

[5] **Addition / Alteration**

The additions / alterations of the equipment stated under [2] b) compared to the equipment already certified comprise of:

- a) An expansion of the series of Pressure measuring devices around the type AX12 HMP331..., this one also authorized for the use in dust explosive areas.
- b) The equipment has another system-unit cover construction.
- c) The electrical connection is executed about a new connection board.

[6] **Test Report**



The explosion proof protection of the equipment stated under [2] b) is documented in the Test Report IB-05-3-384 from 24<sup>th</sup> March 2006. The test documents are component of the Test Report.

[7] **Test result**

IBExU certifies, that the equipment stated under [2] b) fulfills the in Annex II of the RL 94/9/EC fixed Essential Health and Safety Requirements by accordance with EN 50014:1997+A1+A2, EN 50020:2002 and EN 50281-1-1:1998.

The equipment stated under [2] b) fulfills the requirements of explosion protection for electrical equipment in type of protection Intrinsically safety 'ia', Explosion Group IIB, Temperature Class T4 respectively in Dust-type of protection by enclosure with a maximum surface temperature of maximum 85 °C for Equipment Group II, Category 1G respectively 1/2D.

The marking of the equipment stated under [2] b) shall include the following:

 II 1G EEx ia IIB T4  
 II 1/2D IP 6X T 85 °C  
-25 °C ≤ T<sub>a</sub> ≤ +70 °C

[8] **Special conditions**

Besides the special conditions of the equipment certified till now the following is valid in addition:

At use of the HMP 331 in areas of the zone 21 or 22 high energetic loading up mechanisms have to be excluded at the outer surface of the pressure-measuring device (e.g. pneumatic particle transport) at the application.

# IBExU Institut für Sicherheitstechnik GmbH


An-Institut der TU Bergakademie Freiberg

This addition is only valid in connection with the EC-Type Examination Certificate  
IBExU05ATEX1106 X of 03<sup>rd</sup> August 2005.

IBExU Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - 📠 +49 (0) 3731 23650

Authorized for certifications  
- Explosion protection -

By order



(Dr. Lösch)



- Seal -  
(Identification No. 0637)

Freiberg, 28<sup>th</sup> March 2006

Certificates without signature and seal aren't valid.  
Certificates may only be duplicated completely and unchanged.  
In case of dispute, the German text shall prevail.



- [1] **2<sup>nd</sup> Addition to  
EC-TYPE EXAMINATION CERTIFICATE IBExU05ATEX1106 X**  
according to Directive 94/9/EC, Annex III  
- Translation -

- [2] Equipment:  
a) Title corresponding to IBExU05ATEX1106 X:  
**Pressure measuring device**  
type AX12 x|act nn ..., AX12 XMP nn ... and AX12 HMP331...  
b) Supplementary titles: type AX12 LMP380-520 und AX12 XMD...

- [3] Manufacturer: BD SENSORS GmbH

- [4] Address: BD-Sensors-Str. 1  
95199 Thierstein  
Germany

[5] **Addition / Alteration**

The additions / alterations of the equipment mentioned in [2] compared with the equipment already certified concern in detail:

- a) The pressure measuring devices are equipped alternatively with new electronic modules.  
As a result changing the intrinsic safety values of the following:  
effective inner capacity  $C_i = 0 \text{ nF}$   
effective inner inductivity  $L_i = 0 \text{ } \mu\text{H}$   
plus line inductivities  $1 \text{ } \mu\text{H/m}$  and line capacities  $160 \text{ pF/m}$  (cable supplied by the manufacturer)  
The supply connections have an inner capacity of max.  $27 \text{ nF}$  to the housing.
- b) Extension for two new types with another constructive design.
- c) The equipment fulfils also the requirements of the current standard EN 60079-0:2009.

[6] **Test Report**

The proof of the explosion protection of the equipment mentioned in [2] is documented in the Test Report IB-10-3-153 of 8 July 2011. The test documents are part of the Test Report.

[7] **Test result**

IBExU certifies that the equipment mentioned in [2] fulfils the Essential Health and Safety Requirements fixed in Annex II of the Directive 94/9/EC by compliance with EN 60079-0:2009, EN 60079-11:2007, EN 60079-26:2007 and EN 61241-11:2006.

The equipment stated under [2] fulfills the requirements of explosion protection for electrical equipment in type of protection Intrinsic safety 'ia', Explosion Group IIC/IIB, Temperature Class T4 for Equipment Group II, Category 1G, 1/2G or 2G as well as 1D.

The marking of the equipment stated under [2] happen in accordance with a marking table and must include one of the following:

$\text{Ex}$  II 1G Ex ia IIC T4 Ga, II 1/2G Ex ia IIC T4 Ga/Gb, or II 2G Ex ia IIB T4 Gb  
 $\text{Ex}$  II 1D Ex ia IIIC T85 °C Da

**This addition is only valid in combination with the EC-Type Examination Certificate IBExU05ATEX1106 X of 3 August 2005 and their 1<sup>st</sup> Addition.**



**IBExU Institut für Sicherheitstechnik GmbH**  
An-Institut der TU Bergakademie Freiberg

The safety instructions are unchanged and have to be taken into account furthermore.


**IBExU** Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - 📠 +49 (0) 3731 23650

Authorized for certifications  
- Explosion protection -

By order



(Dr. Wagner)



- Seal -  
(Identification No. 0637)

Freiberg, 8 July 2011

Certificates without signature and seal aren't valid.  
Certificates may only be duplicated completely and unchanged.  
In case of dispute, the German text shall prevail.

- [1] **3<sup>rd</sup> Addition to  
EC-TYPE EXAMINATION CERTIFICATE IBExU05ATEX1106 X**  
according to Directive 94/9/EC, Annex III  
- Translation -



- [2] Equipment: **Pressure measuring device**  
Type AX12 x|act nn ..., AX12 XMP nn ..., AX12 HMP331...,  
AX12 LMP380-520 und AX12 XMD...
- [3] Manufacturer: BD SENSORS GmbH
- [4] Address: BD-Sensors-Str. 1  
95199 Thierstein  
Germany

[5] **Addition / Alteration**

The additions / alterations of the equipment mentioned in [2] compared with the equipment already certified concern in detail:

- a) The changing of safety relevant parameter.

$P_i = 680 \text{ mW}$

$I_i = 98 \text{ mA}$

- b) Encapsulation with PU515/PH27 (manufacturer ELCHEMCo spol. S.r.o) as option for rising the resistance against adverse environmental conditons.

- c) The equipment fulfils also the requirements of the current standard EN 60079-0:2012 and EN 60079-11:2012.

The other parameters remain unchanged to previous certificates.

[6] **Test Report**

The proof of the explosion protection of the equipment mentioned in [2] is documented in the Test Report IB-14-3-034 of 18 March 2014. The test documents are part of the Test Report.

[7] **Test result**

IBExU certifies that the equipment mentioned in [2] fulfils the Essential Health and Safety Requirements fixed in Annex II of the Directive 94/9/EC by compliance with EN 60079-0:2012, EN 60079-11:2012 and EN 60079-26:2007.

The equipment mentioned under [2] fulfils the requirements of explosion protection for electrical equipment in type of protection Intrinsic safety 'ia', Explosion Group IIC/IIB, Temperature Class T4 for Equipment Group II, Category 1G, 1/2G or 2G as well as 1D.

The marking of the equipment mentioned under [2] happen in accordance with a marking table and must include one of the following:

$\text{Ex}$  II 1G Ex ia IIC T4 Ga, II 1/2G Ex ia IIC T4 Ga/Gb or II 2G Ex ia IIB T4 Gb  
 $\text{Ex}$  II 1D Ex ia IIIC T85 °C Da

**This addition is only valid in combination with the EC-Type Examination Certificate IBExU05ATEX1106 X of 3 August 2005 and their Additions.**



**IBExU Institut für Sicherheitstechnik GmbH**  
An-Institut der TU Bergakademie Freiberg

The safety instructions are unchanged and have to be taken into account furthermore.


**IBExU** Institut für Sicherheitstechnik GmbH  
Fuchsmühlenweg 7 - 09599 Freiberg, Germany  
☎ +49 (0) 3731 3805-0 - 📠 +49 (0) 3731 23650

Authorized for certifications  
- Explosion protection -

By order



(Dr. Wagner)



- Seal -  
(Identification No. 0637)

Freiberg, 18 March 2014

Certificates without signature and seal aren't valid.  
Certificates may only be duplicated completely and unchanged.  
In case of dispute, the German text shall prevail.